

Information sheet PTS

Mycoplasma (Mg/Ms) antibody detection– VLDIA226

These products are produced by the GD Animal Health and meant to be used in the Proficiency testing scheme.
(Poultry serum – antibodies against Mycoplasma Mg&Ms)

Samples

The sample set consists of 8 coded chicken sera (8 x 250µl) that have to be tested in duplicate (in different test runs) using the method(s) (of choice) your laboratory. The samples should be tested under normal routine test conditions. The samples have been *heat inactivated and freeze-dried (in glass vials)* and should be reconstituted with sterile aquadest.

- 1 Set PTS samples.
1 set consists of 8 lyophilized inactivated poultry serum samples 250µL each, total volume 2.0 ml
- The reconstituted samples should be stored at 2-8 °C and tested within 5 days.
- Sample volumes will be sufficient to perform the required tests.

These products are produced, bottled and sealed at GD Animal Health and meant to be used in the Proficiency testing scheme. No commercial value

All sera are heat inactivated for 30' at 56°C, all other samples are treated with β-Propiolactone.

Reporting your findings

You can use ELISA (please state the manufacturer of the test), Haemagglutination Inhibition (HI) test and/or Rapid or Serum Plate Agglutination (RPA).

- Results from indirect ELISA should be reported as actual titres (when appropriate) or S/P (Serum-To-Positive) values (maximum 3 digits) when appropriate. Please also report the brand of ELISA kit you have used and your interpretation (negative / positive) in the spaces provided. When commercial ELISA's are used, the cut-off as indicated by the manufacturer should be used.
- Blocking ELISA should be reported in % inhibition or S/N ratio.
- HI test results should be reported as log₂ titres and as negative / positive results in the columns provided. Please also report the manufacturer of the antigen.
- RPA test results should be reported as log₂ titres and as negative / positive results in the columns provided. Please also report the manufacturer of the antigen.

Accreditation

The PTS is accredited according to the international standard ISO 17043:2010.